


Oblique Testing with WorldSID, ES2-re, & PMHS

N. Yoganandan
John R. Humm
James Rinaldi
Frank Pintar

Department of Neurosurgery
Milwaukee, WI



Oblique versus Pure Lateral Loading




Modular Scalable Load Wall


- For oblique side impact tests
- Region-specific responses
- Modular load wall design
- Dummy biofidelity issues




Modular, Scalable Load Wall





WorldSID Alignment




WorldSID Alignment





Test Matrix and instrumentation

- Velocities: 3.3, 6.7, and 7.5 m/s
- WorldSID, ES2-re, & PMHS tests
- Repeated testing protocol used
- Triaxial load cells: STAPP plates
- Thorax & abdomen chestbands
- Region-specific ATD deflections
- Oblique and pure lateral output















Results: Chestband-based Data






- Effective peak deflections
- Effective peak angulations
- “IR-TRACC-type” peak deflections
- “Potentiometer-type” deflections
- Thoracic and abdominal regions





Results: Sensor-based Data

- The thoracic and abdominal regions
- Peak deflections from five II IR-TRACC
- Peak deflections from potentiometers









- Summary**
- Region-specific responses
 - > Both ATD: effective in sensing the pure lateral loads
 - > Both ATD: peak deflections in oblique < pure lateral
 - > WorldSID better sensed oblique loading than ES2-re
 - > WorldSID suitable both oblique/pure lateral impacts
 - Current 1D IR-TRACC sensor location
 - > Replicates pure lateral response well
 - > Less than optimal for oblique loading
 - 2D IR-TRACC use and implications
 - Optical sensors may also be used
 - Injury criteria for oblique loading

Thank You

Acknowledgement

This research was supported in part by
US DOT NHTSA Cooperative Agreement
DTNH22-07-H-00173

MEDICAL COLLEGE OF WISCONSIN